

MPC-325 SERIES

MULTI MOTORIZED-MICROMANIPULATOR
CONTROL SYSTEM

OPERATION MANUAL

REV. 3.05 (20120827)



SUTTER INSTRUMENT

ONE DIGITAL DRIVE
NOVATO, CA 94949

VOICE: 415-883-0128 WEB: WWW.SUTTER.COM
FAX: 415-883-0572 EMAIL: INFO@SUTTER.COM



Copyright © 2012 Sutter Instrument Company. All Rights Reserved.

DISCLAIMER

The **MPC-325**-series system consists of one or two MPC-200 controllers, an ROE-200 user control device, and one or more electromechanical micromanipulator devices. The purpose of the system is for the manipulation at the micro level of micropipettes and probes used in conjunction with a microscope. No other use is recommended.

This instrument is designed for use in a laboratory environment. It is not intended, nor should it be used in human experimentation or applied to humans in any way. This is not a medical device.

Do not open or attempt to repair the instrument. High voltages are present and inadvertent movement of the micromanipulator electromechanical could cause injury.

Do not allow unauthorized and/or untrained operative to use this device.

Any misuse will be the sole responsibility of the user/owner and Sutter Instrument Company assumes no implied or inferred liability for direct or consequential damages from this instrument if it is operated or used in any way other than for which it is designed.

SAFETY WARNINGS AND PRECAUTIONS

Electrical



- Operate the MPC-200 using 110-- 240 V AC., 50-60 Hz line voltage. This instrument is designed for use in a laboratory environment that has low electrical noise and mechanical vibration. Surge suppression is recommended at all times. If the system includes two MPC-200 controllers daisy-chained together, surge suppression is strongly recommended.

-  Fuse Replacement: Replace only with the same type and rating:

T2A, 250V, 5 x 20mm, Time Delay fuse (IEC 60127-2, Sheet III)
(Examples: Bussmann GDC-2A. GMC-2A or S506-2-R (RoHS); or
Littelfuse 218 200 or 218 200P (RoHS))

A spare fuse is located in the power input module. Please refer to the fuse-replacement appendix for more details on fuse ratings and for instructions on how to change the fuse.

Avoiding Electrical Shock and Fire-related Injury


-  Always use the grounded power supply cord set provided to connect the system to a grounded outlet (3-prong). This is required to protect you from injury in the event that an electrical hazard occurs.
- Do not disassemble the system. Refer servicing to qualified personnel.
-  To prevent fire or shock hazard do not expose the unit to rain or moisture.

Electromagnetic Interference







To comply with FDA and CE electromagnetic immunity and interference standards; and to reduce the electromagnetic coupling between this and other equipment in your lab always use the type and length of interconnect cables provided with the unit for the interconnection of one or more MP-2x5/M electromechanical devices, host computer via USB interface, (see the Technical Specifications appendix for more details).

Operational

Failure to comply with any of the following precautions may damage this device.

- This instrument is designed for operation in a laboratory environment (Pollution Degree I) that is free from mechanical vibrations, electrical noise and transients.
- This unit is not designed for operation at altitudes above 2000 meters nor was it tested for safety above 2000 meters.
-  **DO NOT CONNECT OR DISCONNECT THE CABLES BETWEEN THE CONTROLLER AND THE MECHANICAL UNITS WHILE POWER IS ON.**

Please allow at least 20 seconds after turning the unit off before disconnecting the mechanical units. Failure to do this may result in damage to the electronics.

- Operate this instrument only according to the instructions included in this manual.
- Do not operate if there is any obvious damage to any part of the instrument.
-  Operate only in a location where there is a free flow of fresh air on all sides. **NEVER ALLOW THE FREE FLOW OF AIR TO BE RESTRICTED.**
-  Do not operate this instrument near flammable materials. The use of any hazardous materials with this instrument is not recommended and if undertaken is done so at the users' own risk.
-  Do not attempt to operate the instrument with the manipulator shipping screws in place - severe motor damage may result.
-  Do not operate if there is any obvious damage to any part of the instrument. Do not attempt to operate the instrument with the manipulator shipping screws in place - severe motor damage may result. When transporting the mechanical manipulator, be sure to install the shipping screws supplied in their correct locations. Failure to do this may result in damage to the motors.
-  Never touch any part of the micromanipulator electromechanical device while it is in operation and moving. Doing so can result in physical injury (e.g., fingers can be caught and pinched between the moving parts of the micromanipulator).
-  As with all microinjection devices, sharp micropipettes can fly out of their holder unexpectedly. Always take precautions to prevent this from happening. Never loosen the micropipette holder chuck when the tubing is pressurized, and never point micropipette


holders at yourself or others. Always wear safety glasses when using sharp glass micropipettes with pressure microinjectors.

-  Do not handle the manipulator mechanical while the power is on, and take care to ensure no cables pass close to the mechanical manipulator.

Other

- Use this instrument only for microinjection purposes in conjunction with the procedures and guidelines in this manual.
- Retain the original packaging for future transport of the instrument.
- Some applications, such as piezo-impact microinjection call for the use of mercury in the micropipette tip. The use of any hazardous materials with any Sutter Instrument's instrument is not recommended and if undertaken is done so at the users' own risk.
- When transporting the mechanical manipulator, be sure to install the shipping screws supplied in their correct locations. Failure to do this may result in damage to the motors.
- This instrument contains no user-serviceable components — do not open the instrument casing. This instrument should be serviced and repaired only by Sutter Instrument or an authorized Sutter Instrument servicing agent.
- Sutter Instrument reserves the right to change specifications without prior notice.
- This device is intended only for research purposes.

Handling Micropipettes

 Failure to comply with any of the following precautions may result in injury to the users of this device as well as those working in the general area near the device.

- The micropipettes used with this instrument are very sharp and relatively fragile. Contact with the pulled micropipette tips, therefore, should be avoided to prevent accidentally impaling yourself.
- Always dispose of micropipettes by placing them into a well-marked, spill-proof “sharps” container.

(This page intentionally left blank.)

TABLE OF CONTENTS

DISCLAIMER	iii
SAFETY WARNINGS AND PRECAUTIONS	iii
Electrical	iii
Avoiding Electrical Shock and Fire-related Injury	iii
Electromagnetic Interference	iv
Operational	iv
Other	v
Handling Micropipettes	v
1. INTRODUCTION	1
1.1 Structure of the MPC-325 Documentation Package	1
1.2 Components of the MPC-325 Series	1
2. MPC-200 MULTI-MANIPULATOR CONTROLLER AND ROE-200 INPUT DEVICE OPERATIONS	3
2.1 Electrical Connections and Initial Operating Instructions	3
2.2 Initial Operating Instructions	3
2.3 Main Controls on the ROE-200	5
2.3.1 White Buttons:	5
2.3.2 Black Selector Switches:	7
2.3.3 Other Controls on the ROE-200	8
2.4 Controls on the MPC-200	10
3. MP-225/M MANIPULATOR MECHANICAL MOUNTING INSTRUCTIONS	15
3.1 Mounting MP-225/M to a Stand or Platform	15
3.2 Setting Headstage/Pipette Angle and Pipette Exchange	16
3.3 Headstage Mounting	17
3.4 Modular Construction	18
3.5 Minimizing Electrical Noise	18
3.6 Instructions for Changing Handedness	18
3.7 Instructions Used in Special Installations Only	19
4. MAINTENANCE	21
APPENDIX A. LIMITED WARRANTY	23
APPENDIX B. ACCESSORIES	25
APPENDIX C. FUSE REPLACEMENT	27
APPENDIX D. TECHNICAL SPECIFICATIONS	29
D.1. MP-225/M	29
D.2. MPC-200 Controller	29
D.3. ROE-200	30
APPENDIX E. EXTERNAL CONTROL COMMAND REFERENCE	31
INDEX	35

TABLE OF FIGURES

Figure 2-1. Rear of MPC-200 controller cabinet.	3
Figure 2-2. Top view of ROE-200.....	6
Figure 2-3. Side view of ROE-200.....	8
Figure 2-4. Configuration switches on rear of MPC-200 controller cabinet.	10
Figure 2-5. Configuration switches (rear of MPC-200 controller).	12
Figure 3-1. Angled side view of MP-225/M showing swing gate and mounting adapter plate.	15
Figure 3-2. Locations of screws used to open swing gate and changing pipette angle.....	16
Figure 3-3. Headstage mounting.	17
Figure 3-4. Using a dovetail for headstage mounting.....	17
Figure 3-5. Rotating the MP-225/M for right and left handedness.....	19
Figure C-1. Rear view of the MPC-200 controller cabinet showing the power entry module and fuse location.....	27

TABLE OF TABLES

Table 2-1. Configuration switch settings for different angles of steepness.....	11
Table D-1. MPC-325 Controller cables and receptacles/connectors.....	30
Table E-1. Complete MPC-200 Remote-Control Command Reference.....	31

1. INTRODUCTION

1.1 Structure of the MPC-325 Documentation Package

The MPC-325-Series is a manipulator system comprised of the MPC-200 controller, the ROE-200 input device and one or more MP-225/M stepper motor manipulators. The manual consists of two parts, “Operations” that describes the functions of the MPC-200 controller and ROE input device and “Setup” that describes how to install the MP-225/M mechanicals.

1.2 Components of the MPC-325 Series

Carefully remove all components from the shipping container. In addition to this manual, the following should be included in each MP-325 system

1. **MPC-325** (a single manipulator system):

- MPC-200 controller
- ROE-200 Rotary Optical Encoder input device
- MP-225/M manipulator mechanical
- Two MP-225 Adapter boxes (adapts 3 wire MP-225/M mechanical to DB-25 cable)
- DB-25 cables (connect adapter boxes to the controller).
- RJ-45 cable (8 conductor) connects the ROE to the controller
- RJ-12 cable (6 conductor) daisy-chains two MPC-200 controllers
- USB cable for computer control of the MPC-325-series system
- Power cable appropriate for your location
- X285210 mounting adapter plates and hardware to attach mechanicals to their mounting surfaces
- X285204 four inch dovetail extensions for mounting headstages
- Dovetail rod clamp
- 2.5mm hex wrench(s) for removing the shipping screws
- 1.5mm hex wrench(s) for adjusting pipette angle

2. **MPC-325-2** (a two-manipulator system). Same as for the MPC-325, with the following exceptions:

- Two MP-225/M manipulator mechanicals
- Two MP-225 Adapter boxes (adapts 3 wire MP-225/M mechanical to DB-25 cable)
- Two DB-25 cables (connect adapter boxes to the controller).
- Two X285210 mounting adapter plates and hardware to attach mechanicals to their mounting surfaces
- Two X285204 four inch dovetail extensions for mounting headstages
- Two dovetail rod clamps

3. **MPC-325-3** (a three-manipulator system). Same as for the MPC-325-2, with the following exceptions:
 - Two MPC-200 controllers
 - Three MP-225/M manipulator mechanicals
 - Three MP-225 Adapter boxes (adapts 3 wire MP-225/M mechanical to DB-25 cable)
 - Three DB-25 cables (connect adapter boxes to the controller).
 - Three X285210 mounting adapter plates and hardware to attach mechanicals to their mounting surfaces
 - Three X285204 four inch dovetail extensions for mounting headstages
 - Three dovetail rod clamps
4. **MPC-325-4** (a four-manipulator system). Same as for the MPC-325-3, with the following exceptions:
 - Four MP-225/M manipulator mechanicals
 - Four MP-225 Adapter boxes (adapts 3 wire MP-225/M mechanical to DB-25 cable)
 - Four DB-25 cables (connect adapter boxes to the controller).
 - Four X285210 mounting adapter plates and hardware to attach mechanicals to their mounting surfaces
 - Four X285204 four inch dovetail extensions for mounting headstages
 - Four dovetail rod clamps

IMPORTANT

Once you have unpacked your MPC-325, remove 6 shipping screws, indicated by the red warning tags, from each MP-225/M. You must remove these screws before operating the manipulators. Save the screws, warning tags, and hex wrench in the event you need to transport either manipulator in the future. Once these screws have been removed, handle the micromanipulators with care. The mechanisms can be damaged if the axes are moved without the screws in place.

2. MPC-200 MULTI-MANIPULATOR CONTROLLER AND ROE-200 INPUT DEVICE OPERATIONS

2.1 Electrical Connections and Initial Operating Instructions

Initially, you may want to simply connect the two manipulators, the controller, and the ROE together and try some gross movements in order to get a feel for the controls and how to make simple movements. It is perfectly acceptable to set the manipulators in the middle of a bench top, make all electrical connections and then observe each unit's movement by eye. Even if you wish to directly install the manipulators in your rig, it is useful to follow the initial setup procedure to learn how to move the units to allow easy access to the mounting screws.

1. Connect the power cord to the power entry module on the back of the MPC-200 controller.




Figure 2-1. Rear of MPC-200 controller cabinet..

2. With the power OFF (front panel switch in the "0" position), connect the ROE-200 input box to the MPC-200 controller using the RJ-45 8-conductor cable. * Use the CONTROLLER output on the back of the ROE and the COMMAND INPUT on the back of the controller.
3. With the power OFF, run a DB-25 cable from each of the two MP-225/M mechanicals to the DB-25 connectors marked "MANIPULATOR A" and "MANIPULATOR B" on the back of the controller. *

*  **CAUTION:** *Never connect or disconnect the ROE or the MP-225/M while the power is on!*

2.2 Initial Operating Instructions

After all connections are made, power up the MPC-325 using the O/I switch on the front of the controller. As it initializes, you will see a start up screen on the ROE-200 that briefly displays the name of the device and the version of the installed firmware. As the power switch is the only control you will need to access on the MPC-200, the controller can ultimately be placed in an out of the way location (e.g., under your bench).

Once the start-up sequence has finished, you will see a display that gives the coordinates of the manipulator. The LED marked 1 will light and the left hand corner of the display shows “Drive A” to indicate that the ROE is ready to operate the MP-225/M connected at the MANIPULATOR A output. Confirm that you get a coordinate display and that you have removed the shipping screws from both manipulators. If you do not get a coordinate display, go to the trouble shooting section at the back of the manual. If you have not yet removed the shipping screws, turn the power off again and remove all shipping screws from both manipulators.

All functions necessary during normal operation are provided by 4 push buttons and two rocker switches on the top of the ROE-200. Other setup functions are done via buttons and DIP switches located on the back of the ROE-200 and DIP switches on the back of the MPC-200 controller.

The three ROE knobs control the three axes of either manipulator (right knob X, left knob Y, and top knob Z (see Page 6)). Turn any one of the three knobs and notice that the corresponding axis moves and the coordinate for that axis changes on the MP225/M connected to the MANIPULATOR A output.

The MPC-200 controller and ROE-200 have a built in Centering function. This is activated by pressing the white “CENTER” button on the back of the ROE. If both MP-225/M manipulators are sitting in a wide-open area, and the shipping screws are removed, press the CENTER button. The ROE-200 display will display the message “PLEASE WAIT MOVE IN PROGRESS” and the first manipulator will center. After the CENTER operation is complete, the manipulator axes will each be at the center of travel and the display will read 12500 for X, Y and Z.

From this location, you can move 12500 microns in each direction on each axis. The unit will stop automatically at each end of travel (00000 or 25000 microns). These ends are determined by firmware. Each axis also has magnetic end of travel switches that are not activated in normal operation. If the magnetic switches are activated, you will see the message EOT (for End Of Travel) on one of the displayed axes.

If you wish, you can easily switch to the second manipulator (connected to the MANIPULATOR B output on the back of the controller). This is done by pressing the Manipulator toggle once. The LED marked 2 will light and the left hand corner of the display will change to “Drive B”. While you are controlling the second manipulator, press CENTER to make sure that this manipulator’s coordinate system is initialized. After centering, you can demonstrate that the manual knobs are now moving this manipulator.

When the MPC-200 controller is first turned on, the speed of movement is at its fastest, coarsest Mode. Movement mode can be finer and slower by changing the black “Mode” toggle switch. As MODE increases from 0, smaller movements are commanded by the same turn of the ROE knob. MODE 5 or 6 is probably what you will use for the final approach to a cell. MODE 0 or “Accelerated Mode” is used for fast movements to move the pipette large distances. In MODE 0, when you turn the ROE knobs slowly, you get relatively slow movement that is useful for final moves to place a pipette near a cell. Conversely, when you make prolonged, rapid turns of the ROE knob, the controller/ROE automatically accelerates to maximum speed to allow for prolonged, long distance movements. This would be most useful for manual pipette exchange.

If you toggle from Drive A to Drive B and back again you will see that the display coordinates and Mode settings are maintained for Drive A while you are using Drive B and vice versa.

The remaining functions of the ROE are explained in the next section.

2.3 Main Controls on the ROE-200

2.3.1 White Buttons:

DIAG/NORM: Pressing the DIAG/NORM button will cause the green LED near the button to light, indicating the MPC-200/ROE-200 is in Diagonal mode. In this mode, rotation of the Z-axis knob produces diagonal movement. A second press will put the manipulator back into Normal mode. When in diagonal mode, the X and Y knobs remain active, allowing you to readjust the X and Y positioning of the pipette as you approach a cell in diagonal mode. Angle of diagonal mode movement is set via DIP switches on the back of the MPC-200 controller (see controller DIP switch setting instructions on Pages 10-12). When using MODE 9 (MODE toggle set to 9), Diagonal mode produces short, quick, impulse-like movement that may be useful in sharp pipette impalements.

When you switch to Diagonal mode, the ROE-200 display is changed from absolute to relative coordinates and the current location is set to 0,0,0. This allows users to invoke relative measurements using the display as a measuring device. A fourth coordinate that gives movement along the diagonal is also added for users who wish to measure the movement of along the axis coaxial with a pipette. When you return to Normal mode, the absolute coordinate system is recovered. The relative coordinate feature can be disabled via DIP switch 2 on back of the ROE-200.




Figure 2-2. Top view of ROE-200.

HOME: When pressed, the manipulator will make a move along a stereotypic path to the location 0,0,0 or "home". Home is the location where you would most likely exchange your pipette and is maximal up on the Z-axis, maximal right on the X-axis (maximal left on a left-handed manipulator) and maximal front on the Y-axis. The stereotypic path of the movement is first along the currently set diagonal until either the X-axis or Z-axis reaches its origin (0). Which one of these occurs first is a function of the diagonal angle and the location at the time HOME is pressed. Once the first limit is reached, the unit will move the two remaining axes simultaneously to their origins (0). The only allowed change in this stereotyped move is that the Y-axis move can be eliminated. This is done via DIP switch 8 on the back of the MPC-200 controller. (See Controller DIP switch setting instructions on Page 12).

WORK POS.: This button has three functions:

1. With the STOP/SET button is held down, a momentary press of WORK POS. makes the current location the "Work Position". A beeper will sound to indicate that the operation is complete and the location has been saved. Typically, this is a location where the pipette tip is under the microscope objective and near the cells or tissue of interest.
2. Once you have defined a Work Position, a momentary press of WORK POS. will cause the manipulator to move to the defined Work Position, providing the manipulator last move was to Home. The move will occur along the predefined path that the manipulator moved to get to Home (described above) but in the opposite direction. This is the reason why Work Position moves **must** follow Home moves; the move to Home defines the return trip. In either case, the movement along the diagonal as you come in and out of the preparation/dish/bath should assure that the pipette tip will not hit anything on the way in or out.
3. When WORK POS. is held down for longer than 2 seconds, the current manipulator is locked so that none of the buttons or the ROE knobs will cause it to move. The lock is released by holding WORK POS. down again. A beep will indicate that the lock is enabled or disabled and the display will indicate the locked state.

STOP/SET: This button has two functions:

1. When held down, STOP/SET" performs a "Set" function in combination with the "WORK POS." key. Think of it as a shift key when held down.
2. A momentary press of STOP/SET during a robotic move (see HOME, WORK POS. and CENTER) will immediately "Stop" the movement. **Think of this as your panic button when you see your pipette headed somewhere that you don't want it to go!**

2.3.2 Black Selector Switches:

MODE: The MODE Selector controls the speed and the relative fineness of movement of the manipulator produced by rotating the ROE knobs. As MODE increases from 0 to 9, movement gets finer and slower. As explained in "INITIAL OPERATING INSTRUCTIONS", MODE 0 is Accelerated Mode. In MODE 0, slow turns of the ROE knob produce medium course moves for moving a pipette under a microscope in the vicinity of a cell. Prolonged, fast turns of the ROE knobs cause the controller to accelerate to top speed for long, imprecise movements for rapid manual positioning of the pipette. The remaining MODES (1-9) produce moves of increasing sensitivity and decreasing speed. In practice, most users will find that

MODE 5 or 6 will provide the necessary dexterity of movement for the final approach to a cell. The current MODE setting is displayed in the upper right of the ROE-200 display.

MANIPULATOR: The MANIPULATOR Selector toggles the active manipulator. In the MPC-325, there is only one active manipulator; therefore, the MANIPULATOR Selector has no function. You can toggle the switch, but no change will occur. With only one manipulator attached, the words "ROE-200" are displayed. If the manipulator is connected to the "MANIPULATOR A" output on the back of the MPC-200 controller LED 1 will be lit. If the manipulator is connected to the "MANIPULATOR B" output on the back of the MPC-200 controller LED 2 will be lit. If you connect an additional MP-225/M or MP-285/M manipulator in the future, the controller will automatically detect the second manipulator, the MANIPULATOR toggle will become active and the display will indicate which manipulator is active with the words "Drive A" or "Drive B".




Figure 2-3. Side view of ROE-200..

2.3.3 Other Controls on the ROE-200

CENTER (round push button on the back of ROE-200): CENTER is an initialization function that is used when the unit is first set up and occasionally during normal operation. **CENTER should only be done in the absence of a pipette as the manipulator makes large robotic movements to its extreme ranges of motion.** To CENTER, press and release the white button on the back of the ROE-200. This will cause a prolonged movement in each axis to the end of travel (EOT) sensors beyond the origin (0,0,0). Once the sensors are found, a short move in the opposite direction is made and this location is defined as (0,0,0). Finally, the unit moves to the location (12500, 12500, 12500), the center of travel of each axis. If the unit is turned off, or STOP/SET is pressed during the running of CENTER, the unit will not be correctly initialized. In this case, it is necessary to cycle the power off and on and run CENTER again to its completion.

DIP Switches (on back of ROE-200): There are four DIP switches on the back of the ROE-200 which govern global and/or ROE settings.

Switch 1: When ON disables all MODES on the MODE Selector except for MODE 0 and 5. Some users may find that they only need Accelerated MODE and a single fine

MODE. This will allow them to more easily switch between the two. Factory default is OFF, enabling all MODES.

Switch 2: When OFF, disables relative coordinates during Diagonal Mode. Factory default is ON, relative display enabled during Diagonal Mode.

Switch 3: When OFF, the MANIPULATOR Selector functions in a cyclical fashion. After reaching the highest number manipulator, a further push of MANIPULATOR cycles the user back to the lowest number manipulator. When DIP switch 3 is set to ON, the selector does not cycle back to the first manipulator. Factory default is OFF, allowing cycling back.

Switch 4: Reserved for future use. **Must be kept ON for proper functioning!**

NOTES

2.4 Controls on the MPC-200

Power Switch: The power switch for the MPC-200 is located on the front panel of the controller. At power up, the microprocessor in the ROE-200 scans the attached equipment and configures the system accordingly. Among the checks/configurations that are made:

1. Determines the number and type of manipulators that are attached. The MPC-200/ROE-200 system is able determine how many and what type of manipulators (MP-285/M or MP-225/M) are connected and to what outputs they are connected. It then sets the current for each output to the correct value for the mechanicals found. If no manipulators are found, the controller will return the message “NO MANIPULATOR DETECTED, PLEASE TURN OFF CONTROLLER AND ATTACH MANIPULATOR”
2. The ROE-200 is capable of connecting to more than one MPC-200 controller. On power up the ROE makes a determination of how many controllers are attached and configures properly. If the power is off on the second controller, the ROE-200 displays a message “PLEASE TURN ON ALL CONTROLLERS, THEN PRESS SET TO START”.

DIP Switches: Two banks of 8 DIP switches are located on the back of the MPC-200 controller. Each bank is assigned to, and configures, one of the two manipulator outputs on the back of the controller (MANIPULATOR A or B). Users familiar with the Sutter Instrument MP-225 controller will find that they have the same function as the configuration DIP switches on the MP-225 ROE. The switches are numbered 1 through 8. In all cases, the 0 or OFF position is opposite the direction of the switch number and the 1 or ON position is in the direction of the switch number and is also indicated by an arrow and the word “ON” next to Switch 1. In order for any new switch settings to take effect, the controller must be powered off and on.

The figure below shows the two banks of switches on the back of the MPC-200 controller.




Figure 2-4. Configuration switches on rear of MPC-200 controller cabinet.


Switches 1, 2, 3 and 4 set the angle of the Diagonal mode movement.

The following table provides the angles that can be used and the DIP switch settings of switches 1, 2, 3 and 4. As indicated in the inset to the left of the table, the angles fall into two different quadrants according to whether the angles are more or less steep than 45 degrees.

Table 2-1. Configuration switch settings for different angles of steepness.



Angle	DIP switch number			
	1	2	3	4
7	1	1	1	1
11	0	1	1	1
14	1	0	1	1
21	0	0	1	1
27	1	1	0	1
29 *	0	1	0	1
35	1	0	0	1
39	0	0	0	1
45	1	1	1	0



Angle	DIP switch number			
	1	2	3	4
39	0	1	1	0
35	1	0	1	0
29	0	0	1	0
27	1	1	0	0
21	0	1	0	0
14	1	0	0	0
11	0	0	0	0

*Factory default near 30 degrees

Switches 5, 6 and 7 set the direction of the movement produced by a clockwise turn (advancing right hand screw) of the ROE knob for each axis.

With the switch set to 0, a clockwise turn of the knob produces a decrement in the display; when the switch is set to 1, a clockwise turn of the knob produces an increment in the display. An increment in the display coincides with movement downward in the Z axis,

movement toward the rear of your setup in the Y axis and movement producing pipette advancement in the X axis.

The factory default is 1,1,1 for switches 5,6 and 7.

Switch number	5	6	7
Corresponding axis	X	Y	Z

Switch 8 determines whether or not the Y-axis is included in HOME and WORK POS. robotic moves. If switch 8 is set to 0, the Y axis is moved to a location where the pipette is towards the user in HOME move and is moved back to whatever Y coordinate was recorded during SET-WORK POS. in the WORK POS. move. If switch 8 is set to 1, the Y axis is not moved (Y position ignored) during the HOME or WORK POS. moves. The factory default for switch 8 is 0; the Y axis will move during HOME and WORK POS. moves.

Remember that the settings on the A switches apply to the MANIPULATOR A output and the settings on the B switches apply to the MANIPULATOR B output. Thus, you can have, for example, different angles of approach on your two manipulators or a different direction of turning to advance the pipette on a left versus a right-handed manipulator.



Figure 2-5. Configuration switches (rear of MPC-200 controller).

(This page intentionally blank.)

3. MP-225/M MANIPULATOR MECHANICAL MOUNTING INSTRUCTIONS

The following sections describe how to mount your MP-225/M to a stand using the mounting adapter plate, how to adjust pipette angle and change pipettes, how to mount different headstages and finally, the modular nature of the mechanical. It is assumed that if you are setting up an MPC-325-2, that you will repeat the setup instructions for two mechanicals. The figure below shows a right-handed MP-225/M. You may have a left-handed unit, especially if you have a two-manipulator system, but the setup is identical.

3.1 Mounting MP-225/M to a Stand or Platform

The MP-225/M mounts to the mounting adapter plate (X285210) using four tapered pegs with locking screws. The figure below shows the location of the front locking setscrews:




Figure 3-1. Angled side view of MP-225/M showing swing gate and mounting adapter plate.

The rear pair of screws is in a similar location in the back of the manipulator. **The manipulator is shipped with the adapter plate in place; it must be removed to mount the manipulator!**

To remove the plate, first loosen the locking screws and then pull the mounting plate straight down. The figure above shows the plate removed from the X axis. Once removed, the mounting adapter plate can be secured to any flat surface carrying 1/4-20 or 10-32 holes on

one-inch centers (such as a Sutter MT-stand or MD series platform) using the supplied hardware. After mounting the adapter plate on your stand or platform, align the pegs on top of the plate with the holes in the manipulator, push the X-axis firmly onto the plate, and retighten the locking hex set screws.

3.2 Setting Headstage/Pipette Angle and Pipette Exchange

Mounted on the front of the Z-axis of the manipulator is the “swing-out gate”. The swing-out gate is the mounting surface for the rotary dovetail that holds various electrophysiological headstages and/or micro tools at defined angles. The swing-out gate is also provides for easy exchange of pipettes during an experiment.

The angle of the rotary dovetail is adjusted by loosening the hex set screw located on the hinge side of the swing-out gate (see figure below). You can set a particular angle using the knife-edge on the dovetail and the scale on the faceplate. After choosing an angle, press the rotary dovetail firmly into the pocket in the swing gate and retighten the screw to fix the angle.




Figure 3-2. Locations of screws used to open swing gate and changing pipette angle.

To change pipettes, loosen the thumbscrew on the swing-out gate (above left). The gate will open allowing the headstage and holder to rotate almost 90 degrees. After replacing your pipette, make sure to close the gate tightly and tighten the thumbscrew securely while holding the gate closed. The thumbscrew is designed to pull the gate closed with tightening. Tightening with thumb and finger is sufficient.

The height of the swing-out gate on the front of the Z-axis is adjustable. To change the position, open the gate and loosen the 4 Phillips-head screws that mount the swing gate. As shipped from the factory, the gate is positioned to allow access to the Z-axis shipping screw holes. You may find it beneficial to move the gate up before you start using your MP-225/M.

3.3 Headstage Mounting

Axon headstages 203B or CV-7 and the Heka EPC-10 headstage have an integral dovetail that slides directly into the rotary dovetail on the MP-225/M. The figure below shows an example of this type of headstage mounted in a left-handed manipulator and in profile (on the right), the location of the Phillips-head screw that secures the headstage dovetail in its mate on the manipulator.




Figure 3-3. Headstage mounting.

Older Axon and Heka headstages mount using the 4-inch dovetail (X285204) and a plastic plate. A typical headstage of this type is shown mounted in a right-handed MP-225/M (right panel). The plastic plate used with the 4-inch dovetail is shown in the left panel of the figure and the holes are indicated to mount common headstages. Additional holes could be easily added to accommodate less common headstage footprints.




Figure 3-4. Using a dovetail for headstage mounting.

Rod mounted headstages and micro tools are accommodated by use of a rod clamp that fits into the dovetail (not shown). All the headstage adapters and mounting hardware are included with the manipulator and are shipped in a zip lock plastic bag.

3.4 Modular Construction

The three axes of the MP-225/M are identical. They are connected to each other with tapered pegs and hex locking screws, four between each axis. The peg and locking screw attachment is identical to the attachment of the mounting adapter plate to the bottom of the manipulator. The manipulators are shipped pre-assembled in either the right or the left-handed configuration. Because of the modular construction, handedness can be easily switched. If you wish to change one of your mechanicals from right handed to left handed or vice versa see “Instructions for Changing Handedness” included later in this section.

The modular construction of the MP-225/M mechanical allows for some flexibility in the connections between axes. In the standard configuration, the axes are stacked X, Y, and Z from bottom to top (see figure on Page 15). However, the 3 axes can be used separately or assembled in non-standard configurations. **If you assemble the manipulator in other configurations, make sure that the axes do not interfere with one another!** For example, if the Z-axis is mounted in a lower location on the right angle that connects it to the Y-axis, it may interfere with the full travel of the Y-axis.

You may have also received one or more accessories for mounting your MP-225/M and/or modifying the headstage mount to the manipulator (i.e. rotating base, microscope stage mount, gantry stand, dovetail extension). Setup of these accessories is normally covered in documentation accompanying the accessory.

If you intend to use the right angle adapter (285300) with your MP-225/M in order to rotate the manipulator 90 degrees, please see “Instructions used in Special Installations Only” near the end of this manual.

3.5 Minimizing Electrical Noise

We are aware of one potential noise source that users coupling their MP-325 with high-gain, high-input impedance, electrophysiological recording amplifiers may experience. Under certain circumstances, the manipulator and/or the drive cable coming from the controller may act as an antenna picking up electric field noise from nearby electrical equipment and bringing it into your setup. Grounding the manipulator will largely eliminate this noise source. Try to attach to one of the silver Phillips-head screws on the side of one or more of the axes. It should be noted that the manipulator produces negligible electrical noise when it is not moving because it is powered by a linear power supply with no AC current present.


3.6 Instructions for Changing Handedness

To switch from Right-handed to Left-handed:


1. Loosen four hex set screws, two on each side of the Y axis, that lock four tapered pins on the top of the X-axis in the holes in the bottom of the Y-axis (the four screws are indicated in the left and right panels below).
2. Pull the X-axis straight down (left panel).
3. Rotate the X-axis 180 degrees in the X-Y plane so that the motor and wire of the X axis are on the left side of the manipulator (middle panel).
4. Reinsert the pins into the holes in the bottom of the Y axis (right panel).
5. Retighten the hex set screws in the sides of the Y-axis.

6. Open the swing gate and remove the four screws that attach it to the Z-axis (swing gate shown open on Page 15).
 7. Rotate the swing gate 180 degrees in the Z-X plane and reattach to the front of the Z-axis.
 8. Remove the thumbscrew from the bottom of the swing gate and install in the similar hole in the top of the swing gate (rotated swing gate shown in right panel).
- To switch from left to right, simply reverse these directions.

RIGHT HAND



X-AXIS ROTATING



LEFT HAND




Figure 3-5. Rotating the MP-225/M for right and left handedness.

3.7 Instructions Used in Special Installations Only

TO INSTALL AND USE THE RIGHT ANGLE ADAPTER (285300)

Open the swing-out gate and remove it from the front of the MP-225/M by removing the four Phillips-head screws. Next, install the right angle adapter on the front of the MP-225/M using the supplied M3-0.5 hex head screws. Finally, install the swing out gate on the right angle adapter using the four Phillips-head screws. With the right angle in place, the manipulator (right handed) can be turned 90 degrees clockwise so that its bulk faces to the right instead of the back of your microscope.

Having made the 90-degree rotation, if you wish to use the automated features and diagonal movement mode of the MP-225/M, two other minor changes must be made:

First, as the X and Y-axes have interchanged with the 90-degree rotation, the wires controlling these two axes must be switched at the connector box.

Second, in order to have the Y axis movement be in the forward direction during the home move, you need to rotate the Y axis 180 degrees with respect to the rest of the manipulator. To reverse the Y (bottom axis in the rotated configuration), locate and loosen the 4 set screws that attach the Y to the X. These are on the sides of the X or middle axis in the stack. When the set screws are loose, you can move the Y off the X, rotate it 180 degrees and reinsert its mounting posts into the correct holes in the X (the holes with set screws). Finally, press the two axes firmly together and retighten the setscrews. Now the Y-axis will move forward during the home move and back during the work position move.

4. MAINTENANCE

Routine cleaning of the MPC-325-series system is required to prevent excessive dust accumulations. This is done by wiping all exterior surfaces with a dry, soft, cotton cloth.

All cables and connections should be periodically inspected to make sure that all connections are made well and that all connectors are well and evenly seated.

CAUTION: The MP-225/M electromechanical is a precision-machined part, mounted on three stepper motor shafts. As such, it DOES NOT REQUIRE LUBRICATION. Attempting to lubricate any part of the electromechanical assembly will void the warranty and may harm one or more of the motors.

(This page intentionally blank.)

APPENDIX A. LIMITED WARRANTY

- Sutter Instrument Company, a division of Sutter Instrument Corporation, limits the warranty on this instrument to repair and replacement of defective components for two years from date of shipment, provided the instrument has been operated in accordance with the instructions outlined in this manual.
- Abuse, misuse, or unauthorized repairs will void this warranty.
- Warranty work will be performed only at the factory.
- The cost of shipment both ways is paid for by Sutter Instrument during the first three months this warranty is in effect, after which the cost is the responsibility of the customer.
- The limited warranty is as stated above and no implied or inferred liability for direct or consequential damages is intended.
- Consumables, PMTs, galvanometers, and Uniblitz^{®1} shutters are exempt from this warranty.
- An extended warranty for up to three additional years can be purchased at the time of ordering, or until the original warranty expires. For pricing and other information, please contact Sutter Instrument.

¹ Uniblitz[®] is a registered trademark of Vincent Associates.

(This page intentionally blank.)

APPENDIX B. ACCESSORIES

285204	4 inch dovetail extension
285210¹	Mounting adapter plate
285RBI	Rotating base for MP-285
225RBI	Rotating base for MP-225
X285300	Right angle adapter
X285305	Z-axis vertical extension
X285310	Z-axis horizontal extension
BR-AW	Rod holding clamp for XenoWorks injectors
MP-ROD	Rod holder
285HEA	Hinged headstage mount

¹ For use with MT- or MD-series stands/platforms, or any surface with 1-inch centered holes.

(This page intentionally blank.)

APPENDIX C. FUSE REPLACEMENT

In the event that the controller fails to power up when the power switch is turned on, check the line power fuses to see if either or both have blown. The fuses are located in the fuse holder on the power entry module on the back of the controller. To remove the fuse holder first unplug the power cord from the power entry module. Press down on the lever that is located just above the fuse holder and below the power receptacle, and pry the holder straight out of the power entry module.

The fuse holder holds two fuses. Both fuses are of the same type and rating. If either fuse is blown, it is recommended that both fuses be replaced.

The type and rating of both fuses are as follows:

**5 x 20 mm glass tube, Time Delay (IEC 60127-2, Sheet III)
T2A, 250V**

**(Examples: Bussmann GDC-2A, GMC-2A or S506-2-R (RoHS); or
Littelfuse 218 200 or 218 200P (RoHS))**




Figure C-1. Rear view of the MPC-200 controller cabinet showing the power entry module and fuse location.

(This page intentionally blank.)

APPENDIX D. TECHNICAL SPECIFICATIONS



D.1. MP-225/M





Resolution	Minimal microstep size is 62.5 nanometers per microstep. Display has single micron resolution.
Speed	3 mm/sec. maximum
Drift	< 1 micron/hr drive mechanism

D.2. MPC-200 Controller

Dimensions (H x W x D):	4 x 15.94 x 11 in (10.16 x 40.49 x 27.94 cm)
Weight:	6.65 lb (6 lb., 10.4 oz.; 3.02 kg)
Electrical:	
Input voltage (Mains)	100 - 240 V, 50/60 Hz
Power consumption	170 Watts maximum per MPC-200 controller (340W max. for an MPC-365-3 or MPC-365-4 system)
Mains fuses (rear of cabinet)	refer to the FUSE REPLACEMENT appendix for details
Cables	(Refer to the following tables for a description of all possible cables.)

Table D-1. MPC-325 Controller cables and receptacles/connectors.

Controller Rear Panel Port Connector/Receptacle	Cable Connector Types	Connects to ...	Cable Type	Cable Max. Length
MPC-200: (Power entry) 3-pin male connector	←3-pin power standard (female) 3-pin male→ (Geographical region dependent)	Mains power source.	10A, 250V, with safety ground plug	3 meters (approx. 10 feet)

Controller Rear Panel Port Connector/Receptacle	Cable Connector Types	Connects to ...	Cable Type	Cable Max. Length
MPC-200 Controller Cabinet: MANIPULATOR A or MANIPULATOR B (each a 25-Pin DSUB female receptacle)	 <p>←DB-25 male DB-25 female→ (Straight-through)</p>	MP-225M (adaptor) (MP-225M Adapter has three RJ45 connectors for three cables that connect to the MP-225M electromechanical)	Minimum of 26 awg stranded wire with 500 Volt.	3 meters (approx. 10 feet)
MPC-200: COMMAND INPUT and ROE-200 CONTROLLER	 <p>←RJ45 male RJ45 female→</p>	ROE-200 User Input Device receptacle named "CONTROLLER"		
MPC-200: EXPANSION	 <p>←RJ11 male RJ11 female→</p>	a second MPC-200 controller (and its EXPANSION receptacle)		
ROE-200: USB USB "A" (Device) female receptacle (full-sized)	 <p>←A connector B connector→</p>	Host computer's USB "B" receptacle (full-sized)	Dielectric separation of circuits. Foil shielding.	3 meters (approx. 10 feet)

D.3. ROE-200

Dimensions:

10 x 6 x 4 in (25.4 x 15.24 x 10.16 cm)

Weight:

3.5 lbs (1.6 kg)

APPENDIX E. EXTERNAL CONTROL COMMAND REFERENCE

This appendix provides a complete list of remote control commands and return values.

Table E-1. Complete MPC-200 Remote-Control Command Reference.

Description		Total Bytes	Byte	Value			Alt-keypad entry	Ctrl-char	ASCII def./-char.	Details
				Dec.	Hex.	Binary				
Change Drive (Manipulator)	Command	3	1	73	49	0100 1001	Alt-0073		I	Command
			2	1 - 4	1 - 4	0000 0001 - 0000 0100	Alt-0001 - Alt-0004	^ A - ^ D	<SOH> - <EOT>	Manipulator number of which to change (1 through 4)
	Return	1	1	13	0D	0000 1101			<CR>	Command completion indicator
Get Current Position	Command	2	1	67	43	0100 0011	Alt-0067		C	Command
	Return	13	1 - 4							X coordinate (-400,000 – 400,000) stored in 4 contiguous bytes as a single “long” (32 bit) signed integer value
			5 - 8							Y coordinate (-400,000 – 400,000) stored in 4 contiguous bytes as a single “long” (32 bit) signed integer value
			9 - 12							Z coordinate (-400,000 – 400,000) stored in 4 contiguous bytes as a single “long” (32 bit) signed integer value

Description		Total Bytes	Byte	Value			Alt-keypad entry	Ctrl-char	ASCII def./-char.	Details	
				Dec.	Hex.	Binary					
			1	13	0D	0000 1101			<CR>	Command completion indicator	
Move to Position	Command	14	1	4D	77	0100 1101	Alt-0077		M	Command	
			2 - 6								X coordinate (-400,000 – 400,000) stored in 4 contiguous bytes as a single “long” (32 bit) signed integer value
			6 - 8								Y coordinate (-400,000 – 400,000) stored in 4 contiguous bytes as a single “long” (32 bit) signed integer value
			10 - 13								Z coordinate (-400,000 – 400,000) stored in 4 contiguous bytes as a single “long” (32 bit) signed integer value
			14	13	0D	0000 1101	Alt-0013	^ M	<CR>	Command terminator	
	Return	1	1	13	0D	0000 1101			<CR>	Command completion indicator	
Move to Home Position	Command	2	1	72	48	0100 1000	Alt-0072		H	Command	
	Return	1	1	13	0D	0000 1101			<CR>	Command completion indicator	
Move to Work	Command	2	1	89	59	0101 1001	Alt-0089		Y	Command	

Description		Total Bytes	Byte	Value			Alt-keypad entry	Ctrl-char	ASCII def./-char.	Details
				Dec.	Hex.	Binary				
Position (as defined by the ROE-200)										
	Return	1	1	13	0D	0000 1101			<CR>	Command completion indicator
Move to Center Position	Command	2	1	78	4E	0100 1110	Alt-0078		N	Command
	Return	1	1	13	0D	0000 1101			<CR>	Command completion indicator
Interrupt Move	Command	2	1	3	03	0000 0111	Alt-0003	^ C	<ETX>	Command
	Return	1	1	13	0D	0000 1101			<CR>	Command completion indicator

NOTE: All coordinates sent to and received from the MPC-200 are in microsteps. To convert microsteps to micrometers, divide the microsteps value by 16 (e.g., 1600 microsteps / 16 = 100 micrometers). To convert micrometers to microsteps, multiply the micrometers value by 16 (e.g., 100 micrometers * 16 = 1600 microsteps).

(This page intentionally blank.)

INDEX

C

Changing Handedness	16
cleaning.....	19
Components	1
controller	
cable specs	27
Controls	5, 9
MPC-200.....	9
DIP Switches.....	9
Power switch	9
ROE-200	5
black selector switches	6, 7
other	
DIP switches	7
other.....	7
CENTER	7
white buttons	5
DIAG/NORM.....	5
HOME.....	6
STOP/SET.....	6
WORK POS.....	6

D

dimensions	
MPC-200 controller	27
ROE-200	28
disclaimer	iii

E

Electrical Connections	3
------------------------------	---

F

fuse	
holder	25
location	25
replacement.....	25
spare.....	25
fuses, replacement	
mains.....	iii, 27

G

glassware	
precautions	v

I

Initial Operating Instructions	3
Installing and using right angle adapter	17

M

mains	
fuses	iii, 27
voltage.....	27
maintenance.....	19
Minimizing Electrical Noise	16
Mounting	
headstage	15
MP-225/M Manipulator Mechanical	13
MP-225/M to a Stand or Platform	13

N

notes	
user.....	34

P

Pipette Exchange.....	14
power entry module.....	25
precautions.....	iii
electrical.....	iii

R

reference	
remote control commands	29

S

safety warnings.....	iii
electrical.....	iii
mains fuse.....	iii
safety warnings & precautions	
operational.....	iv
safety warnings & precautions	iii, iv
Setting Headstage/Pipette Angle	14
Special installations.....	17

T

technical specifications	
MPC-200 controller	
dimensions.....	27
ROE-200 dimensions.....	28
technical specifications.....	27
MPC-200 controller	27
technical specifications	
weight	27
technical specifications	
weight	
MPC-200 controller	27
technical specifications	
electrical.....	27

technical specifications		<i>V</i>	
ROE-200	28	voltage	
technical specifications		mains.....	27
weight	28	<i>W</i>	
technical specifications		warranty	21
weight		weight	27, 28
ROE-200	28	MPC-200 controller	27
		ROE-200	28

NOTES

ADDENDUM

To

ALL OPERATION MANUALS OF MPC-200/ROE-200-BASED MPC-SERIES SYSTEMS

REV. 1.00 – NOVEMBER 28, 2007

As of Version 3.11 (November 12, 2007) of the firmware for the MPC-200 micromanipulator controller and ROE-200 input device, the CENTER routine associated with the white button on the rear of the ROE-200 has been replaced with a CALIBRATE routine that is less likely to break a pipette. Thus, it can be used in the middle of an experiment when you see the message EOT (end of travel) displayed on the ROE-200. Please note that all references to CENTER in the current manual should be replaced with CALIBRATE. Furthermore, the detailed instructions regarding the centering routine in section 2.3.3 should be replaced with the CALIBRATE instructions below.




Figure 1. Location of the CALIBRATE button on the ROE-200.

CALIBRATE is used in two ways. When the unit is first set up, **CALIBRATE** is used to establish the zero location. Then, occasionally, during normal operation, **CALIBRATE** is used to reestablish the zero location. **CALIBRATE** follows a more conservative path than **CENTER**, and can generally be used in the presence of a pipette.

To **CALIBRATE**, press and release the white button on the back of the ROE-200. The manipulator will back away from the current location along the established diagonal (like a **HOME** move), and ultimately move to the end of travel (EOT) sensors, beyond the origin (0,0,0). Once the sensors are found, a short move in the opposite direction is made and this location is defined as (0,0,0). The purpose of **CALIBRATE** is to allow 0,0,0 or **HOME** to be safely reestablished during the course of an experiment without risking damage to the pipette.

If the unit is turned off, or **STOP/SET** is pressed during the running of **CALIBRATE**, the unit will not be correctly initialized. In this case, it is necessary to cycle the power off and on, and then run **CALIBRATE** again to its completion.