

QA File Parameter

Patient Name : ZZZ_RPC, HEAD_AND_NECK
Patient ID : 51515151
Plan Date : 8/7/2016
SSD : 867.1
SDD : 1000.0
Depth : 132.9
Energy : 0
Angle :

Absolute Dose Comparison

Difference (%) : 3.0
Distance (mm) : 2.0
Threshold (%) : 10.0
Meas Uncertainty : Yes
Use Global % : Yes

Summary (Gamma Analysis)

Total Points : 453
Passed : 406
Failed : 47
% Passed : 89.6
*DTA/Gamma is using 3D Mode

Dose Values in cGy


	CAX	Norm	Sel	Max
Set1		46.12	25.27	58.40
Set2	15.95	49.07	28.72	66.40
Set1-Set2		-2.95	-3.45	-8.00
% Diff		-6.02	-7.03	-16.31
DTA(mm)	NA	1.00	NA	
Coords	0,0	-2,0.5	1,1.5	
(y,x) cm				

Notes

1


Reviewed By :

Set1



Y:\imrtqa\VMAT_QA\Patients\Head&Neck\ZZZ RPC...1.txt

Set2



...DCMTPS_Calculated(ZZZ RPC HEAD AND NECK)_A...

Set1-Set2

